



Air pressure correction table

Luftdruckkorrekturtabelle

Luchtdrukcorrectie-tabel

800-900 mbar

t °C	0%	t-tw °C	950 mbar $\hat{=}$ 540 m			Luftdruck/Atmospheric pressure 1000 mbar $\hat{=}$ 111m			1050 mbar $\hat{=}$ minus 301 m		
			Δ 0%	Δ (t-tw) °C	Δ td °C	Δ 0%	Δ (t-tw) °C	Δ td °C	Δ 0%	Δ (t-tw) °C	Δ td °C
10	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	0.8	0.9	0.1	0.2	0.7	0.1	0.1	0.5	0.0	0.1
	80	1.7	1.9	0.2	0.3	1.4	0.1	0.3	1.0	0.1	0.2
	70	2.6	2.9	0.3	0.6	2.2	0.2	0.4	1.5	0.1	0.3
	60	3.5	3.9	0.4	0.9	2.9	0.3	0.7	2.0	0.2	0.5
	50	4.4	4.9	0.5	1.3	3.7	0.4	1.0	2.5	0.2	0.7
	40	5.4	6.0	0.6	1.7	4.5	0.4	1.3	3.1	0.3	0.9
	30	6.4	7.1	0.7	2.5	5.4	0.5	1.9	3.7	0.4	1.3
	20	7.4	8.2	0.8	3.9	6.2	0.6	3.1	4.2	0.4	2.2
	10	8.4	9.4	0.9	7.2	7.1	0.7	5.8	4.8	0.5	4.2
20	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	1.1	0.7	0.1	0.1	0.5	0.1	0.1	0.3	0.0	0.1
	80	2.3	1.3	0.2	0.3	1.0	0.1	0.2	0.7	0.1	0.1
	70	3.5	2.1	0.3	0.5	1.6	0.2	0.3	1.1	0.1	0.2
	60	4.8	2.8	0.4	0.7	2.1	0.3	0.5	1.5	0.2	0.4
	50	6.2	3.6	0.5	1.0	2.7	0.4	0.8	1.9	0.3	0.5
	40	7.6	4.4	0.6	1.5	3.4	0.5	1.2	2.3	0.3	0.8
	30	9.1	5.3	0.8	2.3	4.0	0.6	1.8	2.7	0.4	1.2
	20	10.6	6.2	1.0	3.3	4.7	0.7	2.5	3.2	0.5	1.8
	10	12.3	7.2	1.2	6.2	5.4	0.9	4.9	3.7	0.6	3.6
30	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	1.4	0.5	0.1	0.1	0.3	0.0	0.1	0.2	0.0	0.0
	80	2.9	0.9	0.1	0.2	0.7	0.1	0.2	0.5	0.1	0.1
	70	4.5	1.4	0.2	0.3	1.1	0.2	0.3	0.7	0.1	0.2
	60	6.2	2.0	0.3	0.5	1.5	0.3	0.4	1.0	0.2	0.3
	50	7.9	2.6	0.5	0.8	1.9	0.4	0.6	1.3	0.2	0.4
	40	9.9	3.2	0.6	1.2	2.4	0.5	0.9	1.6	0.3	0.6
	30	12.0	3.9	0.8	1.8	2.9	0.6	1.4	2.0	0.4	1.0
	20	14.2	4.6	1.0	3.0	3.5	0.8	2.3	2.4	0.5	1.6
	10	16.6	5.4	1.3	5.3	4.1	1.0	4.1	2.8	0.7	2.9
40	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	1.7	0.3	0.1	0.1	0.2	0.0	0.0	0.2	0.0	0.0
	80	3.4	0.6	0.1	0.1	0.5	0.1	0.1	0.3	0.1	0.1
	70	5.3	1.0	0.2	0.2	0.7	0.1	0.2	0.5	0.1	0.1
	60	7.4	1.4	0.3	0.4	1.0	0.2	0.3	0.7	0.2	0.2
	50	9.6	1.8	0.4	0.6	1.4	0.3	0.5	0.9	0.2	0.3
	40	12.1	2.2	0.6	0.9	1.7	0.4	0.7	1.2	0.3	0.5
	30	14.8	2.7	0.8	1.4	2.1	0.6	1.1	1.4	0.4	0.7
	20	17.9	3.3	1.0	2.4	2.5	0.8	1.8	1.7	0.5	1.3
	10	21.3	3.9	1.4	4.8	3.0	1.1	3.7	2.0	0.7	2.6
50	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	1.9	0.2	0.0	0.0	0.2	0.0	0.0	0.1	0.0	0.0
	80	3.9	0.4	0.1	0.1	0.3	0.1	0.1	0.2	0.0	0.1
	70	6.1	0.7	0.2	0.2	0.5	0.1	0.1	0.4	0.1	0.1
	60	8.5	0.9	0.2	0.3	0.7	0.2	0.2	0.5	0.1	0.2
	50	11.2	1.2	0.3	0.4	0.9	0.3	0.3	0.6	0.2	0.2
	40	14.2	1.6	0.5	0.7	1.2	0.4	0.5	0.8	0.3	0.4
	30	17.6	1.9	0.7	1.1	1.5	0.5	0.8	1.0	0.4	0.6
	20	21.5	2.4	1.0	1.8	1.8	0.7	1.4	1.2	0.5	1.0
	10	26.1	2.9	1.4	3.9	2.2	1.1	3.0	1.5	0.7	2.1
60	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	2.1	0.1	0.0	0.0	0.1	0.0	0.0	0.1	0.0	0.0
	80	4.4	0.3	0.1	0.1	0.2	0.1	0.1	0.2	0.0	0.0
	70	6.8	0.5	0.1	0.1	0.4	0.1	0.1	0.2	0.1	0.1
	60	9.6	0.7	0.2	0.2	0.5	0.1	0.2	0.3	0.1	0.1
	50	12.6	0.9	0.3	0.3	0.7	0.2	0.3	0.4	0.1	0.2
	40	16.1	1.1	0.4	0.5	0.8	0.3	0.4	0.6	0.2	0.3
	30	20.1	1.4	0.6	0.8	1.0	0.5	0.6	0.7	0.3	0.4
	20	24.9	1.7	0.9	1.4	1.3	0.7	1.1	0.9	0.5	0.7
	10	30.8	2.1	1.4	3.1	1.6	1.0	2.4	1.1	0.7	1.6
70	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	2.3	0.1	0.0	0.0	0.1	0.0	0.0	0.1	0.0	0.0
	80	4.8	0.2	0.1	0.1	0.2	0.0	0.0	0.1	0.0	0.0
	70	7.5	0.3	0.1	0.1	0.3	0.1	0.1	0.2	0.0	0.1
	60	10.6	0.5	0.1	0.2	0.4	0.1	0.1	0.2	0.1	0.1
	50	14.0	0.6	0.2	0.3	0.5	0.2	0.2	0.3	0.1	0.1
	40	17.9	0.8	0.3	0.4	0.6	0.3	0.3	0.4	0.2	0.2
	30	22.6	1.0	0.5	0.6	0.8	0.4	0.5	0.5	0.3	0.3
	20	28.2	1.2	0.8	1.1	0.9	0.6	0.8	0.6	0.4	0.6
	10	35.5	1.6	1.3	2.4	1.2	1.0	1.9	0.8	0.7	1.3

Airned B.V.

E-mail: info@airned.nl
 Website: www.airned.nl

P.O. Box / Postbus 86,
 6580 AB Malden,
 The Netherlands / Niederlande

Telephone / Telefon / Telefoon:
 +31 24 - 35 82 458,

Fax: +31 24 - 35 85 911,



Air pressure correction table

Luftdruckkorrekturtabelle

Luchtdrukcorrectie-tabel

950-1050 mbar

t°C	0%	t-tw °C	950 mbar $\hat{=}$ 540 m			Luftdruck/Atmospheric pressure			1050 mbar $\hat{=}$ minus 301 m		
			Δ 0%	Δ (t-tw)°C	Δ td °C	Δ 0%	Δ (t-tw)°C	Δ td °C	Δ 0%	Δ (t-tw)°C	Δ td °C
10	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	0.8	0.3	0.0	0.0	0.0	0.0	0.0	-0.2	0.0	0.0
	80	1.7	0.5	0.0	0.1	0.1	0.0	0.0	-0.4	0.0	-0.1
	70	2.6	0.8	0.1	0.2	0.1	0.0	0.0	-0.6	-0.1	-0.1
	60	3.5	1.1	0.1	0.2	0.1	0.0	0.0	-0.8	-0.1	-0.2
	50	4.4	1.3	0.1	0.4	0.2	0.0	0.0	-1.0	-0.1	-0.3
	40	5.4	1.6	0.2	0.5	0.2	0.0	0.1	-1.3	-0.1	-0.4
	30	6.4	1.9	0.2	0.7	0.2	0.0	0.1	-1.5	-0.2	-0.6
	20	7.4	2.3	0.2	1.2	0.3	0.0	0.1	-1.7	-0.2	-1.0
	10	8.4	2.6	0.3	2.4	0.3	0.0	0.3	-2.0	-0.2	-2.3
20	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	1.1	0.2	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	0.0
	80	2.3	0.4	0.0	0.1	0.0	0.0	0.0	-0.3	0.0	-0.1
	70	3.5	0.6	0.1	0.1	0.1	0.0	0.0	-0.4	0.0	-0.1
	60	4.8	0.8	0.1	0.2	0.1	0.0	0.0	-0.6	-0.1	-0.1
	50	6.2	1.0	0.1	0.3	0.1	0.0	0.0	-0.8	-0.1	-0.2
	40	7.6	1.2	0.2	0.4	0.1	0.0	0.0	-0.9	-0.1	-0.3
	30	9.1	1.5	0.2	0.7	0.2	0.0	0.1	-1.1	-0.2	-0.5
	20	10.6	1.7	0.3	1.0	0.2	0.0	0.1	-1.3	-0.2	-0.8
	10	12.3	2.0	0.3	2.0	0.2	0.0	0.2	-1.5	-0.3	-1.8
30	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	1.4	0.1	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	0.0
	80	2.9	0.3	0.0	0.1	0.0	0.0	0.0	-0.2	0.0	0.0
	70	4.5	0.4	0.1	0.1	0.0	0.0	0.0	-0.3	0.0	-0.1
	60	6.2	0.5	0.1	0.1	0.1	0.0	0.0	-0.4	-0.1	-0.1
	50	7.9	0.7	0.1	0.2	0.1	0.0	0.0	-0.5	-0.1	-0.2
	40	9.9	0.9	0.2	0.3	0.1	0.0	0.0	-0.7	-0.1	-0.3
	30	12.0	1.1	0.2	0.5	0.1	0.0	0.1	-0.8	-0.2	-0.4
	20	14.2	1.3	0.3	0.9	0.1	0.0	0.1	-1.0	-0.2	-0.7
	10	16.6	1.5	0.4	1.6	0.2	0.0	0.2	-1.1	-0.3	-1.4
40	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	1.7	0.1	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	0.0
	80	3.4	0.2	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	0.0
	70	5.3	0.3	0.1	0.1	0.0	0.0	0.0	-0.2	0.0	-0.1
	60	7.4	0.4	0.1	0.1	0.0	0.0	0.0	-0.3	-0.1	-0.1
	50	9.6	0.5	0.1	0.2	0.1	0.0	0.0	-0.4	-0.1	-0.1
	40	12.1	0.6	0.2	0.3	0.1	0.0	0.0	-0.5	-0.1	-0.2
	30	14.8	0.8	0.2	0.4	0.1	0.0	0.0	-0.6	-0.2	-0.3
	20	17.9	0.9	0.3	0.7	0.1	0.0	0.1	-0.7	-0.2	-0.5
	10	21.3	1.1	0.4	1.5	0.1	0.0	0.2	-0.8	-0.3	-1.2
50	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	1.9	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	80	3.9	0.1	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	0.0
	70	6.1	0.2	0.0	0.1	0.0	0.0	0.0	-0.1	0.0	0.0
	60	8.5	0.3	0.1	0.1	0.0	0.0	0.0	-0.2	0.0	-0.1
	50	11.2	0.3	0.1	0.1	0.0	0.0	0.0	-0.3	-0.1	-0.1
	40	14.2	0.4	0.1	0.2	0.1	0.0	0.0	-0.3	-0.1	-0.1
	30	17.6	0.5	0.2	0.3	0.1	0.0	0.0	-0.4	-0.1	-0.2
	20	21.5	0.7	0.3	0.5	0.1	0.0	0.1	-0.5	-0.2	-0.4
	10	26.1	0.8	0.4	1.1	0.1	0.0	0.1	-0.6	-0.3	-0.9
60	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	2.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	80	4.4	0.1	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	0.0
	70	6.8	0.1	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	0.0
	60	9.6	0.2	0.1	0.1	0.0	0.0	0.0	-0.1	0.0	0.0
	50	12.6	0.2	0.1	0.1	0.0	0.0	0.0	-0.2	-0.1	-0.1
	40	16.1	0.3	0.1	0.1	0.0	0.0	0.0	-0.2	-0.1	-0.1
	30	20.1	0.4	0.2	0.2	0.0	0.0	0.0	-0.3	-0.1	-0.2
	20	24.9	0.5	0.2	0.4	0.1	0.0	0.0	-0.4	-0.2	-0.3
	10	30.8	0.6	0.4	0.9	0.1	0.0	0.1	-0.4	-0.3	-0.7
70	100	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	90	2.3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	80	4.8	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
	70	7.5	0.1	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	0.0
	60	10.6	0.1	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	0.0
	50	14.0	0.2	0.1	0.1	0.0	0.0	0.0	-0.1	0.0	-0.1
	40	17.9	0.2	0.1	0.1	0.0	0.0	0.0	-0.2	-0.1	-0.1
	30	22.6	0.3	0.1	0.2	0.0	0.0	0.0	-0.2	-0.1	-0.1
	20	28.2	0.3	0.2	0.3	0.0	0.0	0.0	-0.3	-0.2	-0.2
	10	35.5	0.4	0.4	0.7	0.0	0.0	0.1	-0.3	-0.3	-0.6

Airned B.V.

E-mail: info@airned.nl
 Website: www.airned.nl

P.O. Box / Postbus 86,
 6580 AB Malden,
 The Netherlands / Niederlande

Telephone / Telefon / Telefoon:
 +31 24 - 35 82 458,

Fax: +31 24 - 35 85 911,